



**Bihar Medical Services & Infrastructure Corporation  
Limited**  
4<sup>th</sup> floor State Building Construction Corporation  
Limited. Hospital Road, Shastri Nagar, Patna  
800023  
Phone/Fax: +917008050665, + 919471009193

**CORRIGENDUM-I**

**Re-Tender for rate contract and supply of Salt Testing Kit for different healthcare facilities of State of Bihar for the year 2019-21.**

**Notice Inviting Tender Ref No.:-BMSIC/MEDICALDEVICE/CONSUMABLE/19-02 Dated:22-08-2019**  
(Only through E- Tender on website: -[www.eproc.bihar.gov.in](http://www.eproc.bihar.gov.in))

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) invites E-Bids from the interested parties for Tender No.- BMSIC/MEDICAL DEVICE/CONSUMABLE/19-02 entitled as **"Re-Tender for rate contract and supply of Salt Testing Kit for different healthcare facilities of State of Bihar"**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website [www.eproc.bihar.gov.in](http://www.eproc.bihar.gov.in). Kindly refer to the Minutes of the Technical Specification Committee Meeting (including Annexure- A & B), dated- 09<sup>th</sup> August 2019 attached with this Corrigendum in which certain amendments are made. Also, in order to give sufficient time and to ensure wider participation of the bidders the Tender Schedule is being revised as follows.

**Revised Tender Schedule**

Tender Reference No.	<b>BMSIC/MEDICAL DEVICE/CONSUMABLE/19-02</b>
Date and time for downloading of bid document	<b>Upto 03<sup>rd</sup>September 2019 till 1500 Hrs.</b>
Last date and time of submission of online bids	<b>Upto 04<sup>th</sup>September 2019 till 1800 Hrs.</b>
Last date and time for submission of original bid documents with EMD and Tender Fees	<b>06<sup>th</sup>September 2019 till 14:00 Hrs.</b>
Date, Time and Place of opening of Technical Bid	<b>06<sup>th</sup>September 2019 (at 15:00 Hrs.) on the website of <a href="http://www.eproc.bihar.gov.in">www.eproc.bihar.gov.in</a> in the office of BMSICL</b>
Date and time of opening of Financial Bids	<b>To be announced later on <a href="http://www.bmsicl.gov.in">www.bmsicl.gov.in</a> and <a href="http://www.eproc.bihar.gov.in">www.eproc.bihar.gov.in</a></b>
Validity of Tender	<b>180 Days</b>
Cost of the tender document	<b>Rs. 10000/- (Ten Thousand only) Non-refundable.</b>
Bid Processing Fee	<b>Rs 1180/- (One thousand one hundred eighty only) Non-refundable.</b>

**Sd/-  
CGM (Supply Chain)  
BMSICL**

**Note- I** The Specifications of **Salt Testing kit**, mentioned in the price bid have been amended and read as per the specification mentioned in the Annexure- A of the Minutes of the Technical Specification Committee Meeting. Please refer to the Minutes of the Technical Specification Committee Meeting and revised product list in the price Bid before submission of the Bid. Those bidders who have already submitted their bids should also have to Re-submit their bids accordingly

**Note- II** The bidders have to submit the Valid Pollution Control Clearance Certificate in accordance with Water [Prevention and control of Pollution] Act, 1974 & Air [Prevention and control of Pollution] Act, 1981 and Hazardous Wastes (Management, Handling & Trans Boundary Movement) Rules 2008 (Self Attested Copy of Certificate to be enclosed).

**Note- III** The bidders have to provide the detailed address of the Carrying and forwarding Agent/ Sales Depot point/ Dealership point/ Stockist point/ Distributor point in the State of Bihar including the Mobile Number so that Market survey can be done (if required). If your firm is not marketing the quoted products in State of Bihar and doesn't have any Carrying & Forwarding agent/ Sales Depot/ Dealership/ Stockist/ Distributorship then the bidder has to provide a self-declaration on the Firms letter head stating the same.

**Note-IV** In case of any query/doubt one can contact DGM(Drugs) on Mobile No- 7008050665 or Manager(Drugs) on Mobile No- 9471009193 or may contact through e-mail on contact2bmsicl@gmail.com.

**Annexed-:** Minutes of the Technical Specification Committee Meeting (including Annexure- A & B), dated- 09<sup>th</sup> August 2019.

**Sd/-**  
**CGM (Supply Chain)**  
**BMSICL**

**Minutes of the Technical Specification Committee Meeting for Finalization of Technical Specification of Salt Testing Kit floated vide tender no.- BMSIC/MEDICAL DEVICES/CONSUMABLE /19-02 (Tender for rate contract & supply of Salt Testing Kit for different healthcare facilities of State of Bihar for the year 2019-21.**

**Date: 09<sup>th</sup> August 2019**

**Venue: BMSICL Conference Hall.**

The Technical Specification Committee Meeting for finalization of the specification of the products floated vide tender no.- BMSIC/ MEDICAL DEVICES/CONSUMABLE/19-02 was held on 09<sup>th</sup> August 2019 under the Chairmanship of Dr. R. D. Ranjan, Director in Chief, Department of Health, GoB. The other committee members present in the meeting were:

1. Sh. Rajani Kant, CGM (Supply Chain), BMSICL.
2. Dr. Rajan Kumar, Department of Medicine, PMCH.
3. Dr. A.K Shahi, SPO, NIDDCP, SHS, Patna.
4. Sh. Manish Ranjan, Assistant Director, Drugs, SHS, Patna.
5. Sh. Sunil Kumar Singh, GM (Procurement), BMSICL.
6. Dr. Biswaprakash Pradhan, DGM (Drugs), BMSICL.
7. Sh. Ravi Kiran Rakesh, Manager (Drugs), BMSICL.

During and after pre-bid meeting held on 07-08-2019, various suggestion from prospective Bidders were received. All the received suggestion and technical aspects were discussed in today's meeting. After due deliberations on all the aspects, it was unanimously decided to recommend to make certain amendments and clarifications as given in annexure-A and Annexure- B of this minutes of the meeting.

**The meeting was thus concluded with a vote of thanks.**

**Sd/-**  
**(Ravi Kiran Rakesh)**  
Manager (Drugs), BMSICL, Patna.

**Sd/-**  
**(Dr. Biswaprakash Pradhan)**  
DGM (Drugs), BMSICL, Patna.

**Sd/-**  
**(Sunil Kumar Singh)**  
GM (Logistics), BMSICL, Patna

**Sd/-**  
**(Manish Ranjan)**  
Assistant Director, Drug  
SHS, Patna.

**Sd/-**  
**(Dr. A.K Shahi)**  
SPO, NIDDCP, SHS, Patna

**Sd/-**  
**(Dr. Rajan Kumar)**  
Department of Medicine, PMCH

**Sd/-**  
**(Sh. Rajani Kant)**  
CGM (Supply Chain), BMSICL, Patna.

**Sd/-**  
**(Dr. R.D. Ranjan)**  
Director in Chief, Health Dept. Govt. of Bihar.

### Annexure- A

TENDER NO. BMSIC/MEDICAL DEVICES/CONSUMABLES/19-02					
TENDERED PRODUCT LIST AFTER MINOR AMMENDMENT					
S.N.	Name of the Product	Specification / Strength	Dosage Form	Pack Size	Estimated Quantity for 2019-2021
1	Salt Testing Kit	<p><b>1.</b> The salt testing kit should be ready in use, liquid form. Each salt testing kit should have 20 ml testing solution or testing capacity of 75-100 samples. Supply should be in plastic screwed cap vial and able to dispense one drop at the time for use at community level. It can be stored at ambient temperature and relative humidity (20-90%±10%) in any part of the country.</p> <p><b>2. The kit should be able to differentiate: -</b>  Salt with Iodine.  Salt with inadequate iodine in the range of 05 to less than 15ppm.  Salt with adequate levels of Iodine 15ppm and above.</p> <p><b>3.</b> The kits should be able to detect iodine levels in the salt from various sources and characteristics e.g.- salts that are alkaline/acidic in nature and with varying sodium chloride content in the country.</p> <p><b>4.</b> The test kit should have been evaluated and validated by at least one International agencies like WHO, UNICEF, MI,, and / or National level Laboratories such as National Institute of Nutrition, Hyderabad, National Centre for Disease Control, Delhi; All India Institute of Medical Sciences , New Delhi; All India</p>	KIT	<p><b>Pack Size</b>  <b>Salt Testing Kit, with 20ml plastic screwed cap bottle with one spoon, one sample holder and Iodine test colour chart.</b>  <b>Note: There is no change in the specification/ strength.</b></p>	337088

		<p>Institute of Hygiene &amp; Public Health , Kolkata;  Central Food Technological Research Institute , Mysore;  Indian Council of medical Research &amp; Council of Scientific and Industrial Research Laboratories . The validation should include test for the quality, packaging, ready to use testing (drop by drop), stability at various places, shelf life under sealed condition as well as open condition as all these parameters are interlinked. The testing Laboratory should submit detailed report about all the test parameters including how they vary under different field conditions.</p> <p><b>5.</b> The offered Manufacturers/Bidders should have manufacturing and marketing experience minimum of 2 Years and should be supported by documentary evidence.</p> <p><b>6.</b> The shelf life of the salt testing kit should be at least one year and when the vial is opened, it should not be less than 4-6 months.</p> <p><b>7.</b> Pack Size: Each salt testing Kit should be independently packed and not more than 10 kits in a bigger package, for the purpose of ease of transportation/distribution.</p> <p><b>8.</b> Bidders are required to submit documentary proof in support of above quoted specifications and requirements along with their bids.</p> <p><b>9.</b> Bidders are also required to submit the three packets having ten kits each of independent packing as per technical specifications at S.N.- 1 to 6 of salt testing kit as samples along with their bids.</p>			
--	--	--	--	--	--

--	--	--	--	--	--

ANNEXURE-B				
BMSIC/MEDICAL DEVICES/CONSUMABLE/19-02, Clarification in the light of queries/suggestions received from various firms during and after Pre-Bid Meeting held on 07/08/2019				
S.N.	Name of the Firm	Bidders Queries/ Suggestions/ Changes requested	Present Clause	Clarification/Amendment
1	M/s PLASTI SURGE INDUSTRIES PVT LTD	<p>1. 10 Ml testing solutions with Reagent A and Reagent B having testing capacity of 75-100 samples</p> <p>2. Instruction sheet in local language shall be provided in at the time of supply, however with sample kit can provide in English language in absence of art work.</p> <p>3. No separate colour chart needed may provide on individual Box with detailed instruction to use.</p> <p>4. Primary pack of 10 Kits of Individual packing.</p>	<p><b>Pack Size</b></p> <p>1. 2-3 Ampoules of 10 ml each containing test solution.</p> <p>2. In some cases, 1 red ampoule containing recheck solution.</p> <p>3. A detailed instruction sheet in the local language.</p> <p>4. A colour chart with circular colour spots.</p> <p>5. One small Container.</p> <p>6. All fitting into a handy pocket size kit.</p>	<p><b><u>Pack Size</u></b></p> <p><b>Salt Testing Kit, with 20ml plastic screwed cap bottle with one spoon, one sample holder and Iodine test colour chart.</b></p> <p><b>Note: There is no change in the specification/ strength.</b></p>

	<p>We are offering Non Drug products which covered under MSME Registration. Moreover, industries department does not use to issue Market Standing Certificate and Production Capacity Certificate separately as like Drug Authorities but already mentioned capacity in MSME under various heads. Hence it is requested to amend the Market Standing to be certified by CA.</p>	<p><b>Clause 3(h):-</b> Bidder must have Market Standing Certificate (in India) of minimum three years issued by the concerned Licensing Authority from Drugs Control Department for the quoted product (where ever applicable). Self-attested copies are to be submitted.</p>	<p><b>Clause 3(h):-</b> <b>Submission of Market Standing Certificate (in India) of minimum three years issued by the concerned Licensing Authority from Drugs Control Department for the quoted product is <u>not mandatory</u> but the bidder has to justify the Market Standing Criteria by providing the Self-attested copies of Sales Invoices for last three years.</b></p>
	<p>We are offering Non Drug products which covered under MSME Registration. WHO-GMP/GMP as per M/COPP is not Applicable.</p>	<p><b>Clause 3(j) WHO-GMP/GMP (Good Manufacturing Practice)</b> as per revised Schedule-‘M’/COPP Certificate of the manufacturing unit issued by the Licensing Authority/ Drugs Control Department (where ever applicable). The GMP certificate must not be older than one year from the last date of submission of tender. Self-attested copies are to be submitted.</p>	<p><b>Clause 3(j) WHO-GMP/GMP (Good Manufacturing Practice)</b> as per revised Schedule-‘M’/COPP Certificate of the manufacturing unit issued by the Licensing Authority/ Drugs Control Department (where ever applicable) for the quoted product is <u>not mandatory</u>.</p>

2	M/s GLOBAL BIO-SCIENCES	<p>Kindly note that most of the leading manufacturers of salt testing kit who are into single product manufacturing and supplying to multiple state corporations have a turnover of less than 5 crores annually. The manufacturers who have more than 5 crore annual turnover usually manufacture several other products. We are currently supplying to the state corporation like West Bengal, Jharkhand, Chattisgarh, Rajasthan, Kerala, Uttar Pradesh. We hope you will change the turnover for this product to Rs. 1.5 Crore Average 3 years to allow multiple bids and competitive rates like other corporation.</p>	<p><b>Clause 3(1)</b> Copies of the Audited Balance Sheet and Profit and Loss statement showing details of their annual average turnover not less than 5 Crores for any three of the last four consecutive financial years (Auditor/CA certificate of turnover will not be accepted). Self-attested copies are to be submitted.</p>	<p><b>Clause 3(1) Copies of the Audited Balance Sheet and Profit and Loss statement showing details of their annual average turnover not less than <u>1.5 Crores</u> for any three of the last four consecutive financial years (Auditor/CA certificate of turnover will not be accepted). Self-attested copies are to be submitted.</b></p>
---	-------------------------	---	---	--

Sd/-  
CGM (Supply Chain)  
BMSICL