



Bihar Medical Services & Infrastructure Corporation Limited, 4th floor, Bihar State Building Construction Corporation Limited, Hospital Road, Shastri Nagar, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of Medical Equipment for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2021-22/ME-252. During and after Pre-bid meeting various suggestions were received from different prospective bidders regarding amendment in technical specification of equipment which were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:

Tender Reference No.	BMSICL/2021-22/ME-252
Date and time for downloading of bid document	Up to 12th February 2022 till 17:00 Hrs.
Last date and time of submission of online bids	14th February 2022 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document.	15th February 2022 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	15th February 2022 (at 15:00 Hrs.) on the website of www.eproc.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on www.eproc.bihar.gov.in

Note:- Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I		
SL. No	Technical Specification Before Amendment	Technical Specification After Amendment
Syringe Pump		
1	Microprocessor controlled pump capable of propulsion of fluids accurately.	No Change
2	Syringe compatibility: The pump should work with different brands of syringes and is able to accept syringes with volumes of 05 ml – 50 ml.	No Change
3	There should be automatic detection of syringe size.	No Change
4	It should be equipped for detecting correct fixing of syringe.	No Change
5	Flow rate should be adjustable from 0.1 ml/hr to 1000 ml/hr.	Flow rate should be adjustable from 0.1 ml/hr to 1000 ml/hr. or more
6	Flow rate should be adjustable in increments of 0.1 ml/hr.	No Change
7	The accuracy of flow rate should be $\pm 2\%$.	No Change
8	The flow rate should be displayed in ml/ hr.	No Change
9	It should be able to deliver bolus dose in automatic & manual mode.	No Change
10	Internal Rechargeable battery backup should 4 to 6 hrs.	No Change
11	Pump should have colour/Monochrome display with screen size 3 inch or more.	Pump should have colour/Monochrome display with screen size 2.4 inch or more.
12	The following audio and visual alarms should be incorporated: a) Main changeover to battery indication b) Occlusion pressure alarm c) Near empty syringe d) Low battery e) Standby alarm.	No Change
13	The pump should be waterproof so that fluid should not enter inside the pump in case of accidental spillage and necessary Certificate to this effect should be submitted or presented along with the bid.	No Change
14	The syringe pumps should be capable of standalone functioning as well as being fixed on a frame/platform/stand.	No Change
15	US FDA / European CE (issued by notified body) approved model should be offered.	No Change
16	Power Supply 220VAC +/- 10 %, 50Hz fitted with Indian plug.	No Change