



Bihar Medical Services & Infrastructure Corporation Limited, 4th floor, Bihar State Building Construction Corporation Limited, Hospital Road, Shastri Nagar, Patna 800023, Phone/Fax: +91612 2283287,+ 91612 2283288

Corrigendum-I

Bihar Medical Services and Infrastructure Corporation Limited (BMSICL) had invited E-Bids from the interested parties for the procurement, rate contract and the supply of Crash Cart for different Govt. Institutions of Bihar vide Notice Inviting Tender No.- BMSICL/2022-23/ME-294. During and after Pre-bid meeting for the above said tender, no suggestions regarding amendment in technical specifications of the equipment received from different prospective bidders. The technical specification of the equipment were discussed and deliberated on by the experts, who after due deliberation recommended certain amendments in the technical specification of the equipment, which are annexed as Annexure-I of this corrigendum. In order to facilitate maximum participation of bidders the tender schedule is being revised as follows:

Tender Reference No.	BMSICL/2022-23/ME-294
Date and time for downloading of bid document	Up to 09th November 2022 till 17:00 Hrs.
Last date and time of submission of online bids	12th November 2022 till 17:00 Hrs.
Last date and time of submission of original documents of EMD, Tender Fee and Document	15th November 2022 till 14:00 Hrs.
Date, Time and Place of opening of Technical Bid	15th November 2022 (at 15:00 Hrs.) on the website of www.eproc.bihar.gov.in in the office of BMSICL
Date and time of opening of financial Bids	To be announced later on www.eproc.bihar.gov.in

Note:-

- 1. Bidders are advised to refer to the Annexure-I of this corrigendum before submission of bid.**
- 2. Those who have submitted their bids are requested to re-submit their bids in accordance with this corrigendum.**

Annexed:- as above

**Sd/-
GM (Procurement)
BMSICL**

Annexure-I		
Sl.No.	Technical Specification as per tender	Amendment
Name of Equipment: - Mortuary Chamber (Two Body)		
1	Storage temperature: 2-5° C.	No Change
2	Solid state digital controller, vapour proof light with switch.	No Change
3	Electronic Safety system which operates whenever there is fluctuation/ deviation from set temperature operates even in case of power failure by audio/ visual alarm.	No Change
4	Temperature variation adjustment along with safety thermostat present Construction.	No Change
5	All stainless steel 304 which include inner, front door and carriage. Carriage frame to be have smooth and noiseless movement.	No Change
6	Base should be 2- 2.5 feet from the ground level & 304 stainless steel.	No Change
7	Hermetically sealed compressor serves refrigeration unit, to maintain temperature range of 2° C to 5° C Unit is fitted with Digital Temperature Indicator-cum-controller.	No Change
8	Automatic locking facility of carriage for storage when returned to close position.	No Change
9	Air circulation Forced Refrigerant.	No Change
10	PUF insulation on all sides.	No Change
11	Maximum permitted load per shelf – 155 kg and of Body height 6.5 Feet.	No Change
12	US FDA / European CE /BIS/CE Approved model should be offered.	No Change
13	Voltage – 220VAC ±10%,	No Change
14	Suitable stabilizer for smooth functioning of equipment.	Suitable stabilizer for smooth functioning of equipment. (Automated Type)